//AXIS

TORQUE TESTER FSA

- torque measurement in both directions
- → high speed of measurement up to 1000 samples per second
- ➔ peak mode
- storing measurement results into memory
- USB and RS232C serial interface
- statistics, graphs, comparing to thresholds function



Cap torque testers FSA series are designed for measuring torque during opening jar, bottles etc. Measurement processor enables to register maximal force torque needed for example to open a jar. Tester is equipped with a wide range clamp to measure variety of objects.



ext. calibration

results memory



printout configuration

force chart

date and time



_____ kg | Ibs

thresholds/seletion

units selection

multilanguage



ain farca

max force

min force

statistics

Principle of operation



Wide range of adjustable diameters




Easy fixing



Simple measurement





Technical data

Model	FSA2	FSA5	FSA10	
Maximum force measured	2Nm (~0,2kgfi	m) 5Nm (~0,5kgfm)	10Nm (~1kgfm)	
Reading graduation	0,001Nm	0,001Nm	0,01Nm	
Accuracy		±0,5% F.S.		
Maximum overload		150% F.S.		
Operational temperature		-10 ÷ +40°C		
Units	Nm, Ncm, kgfm, Ibfin, gfm			
Display	grafical LCD			
Measurement speed	max 1000 measurements/s			
Measuring functions	maximal value measurement, serial measurement, measurement in time			
Measurements memory	RAM 6 400 measurements, microSD card slot; only FC: optional RAM memory expansion to 72000 measurements or 26000 with date+time			
Interfaces	RS232C, USB, clock; options: trigger IN, transoptor OUT (for thr function)			
Software	AXIS FM program			
Power supply	internal accumulators NiMH: 4x1,2V 2700mAh + supply: ~230V 50Hz / 12V; 1,2A			
Working time with accumu	Ilator	~30h (~25h with backlighting)		
Range of diameters	ø20÷ ø180mm			
Dimensions	345x236x160mm			
Weight	5,1kg			

AXIS Sp. z o.o. 80-125 Gdańsk, ul.Kartuska 375B Poland, tel. +48 58 320 63 01...03, fax 320 63 00 www.axis.pl e-mail: axis@axis.pl